

CONTENTS

CHAPTER 1. BASIC RADIATION PHYSICS	1
1.1. INTRODUCTION	1
1.1.1. Fundamental physical constants (rounded off to four significant figures)	1
1.1.2. Important derived physical constants and relationships ..	1
1.1.3. Physical quantities and units	3
1.1.4. Classification of forces in nature	4
1.1.5. Classification of fundamental particles	4
1.1.6. Classification of radiation	5
1.1.7. Classification of ionizing photon radiation	6
1.1.8. Einstein's relativistic mass, energy and momentum relationships	6
1.1.9. Radiation quantities and units	7
1.2. ATOMIC AND NUCLEAR STRUCTURE	7
1.2.1. Basic definitions for atomic structure	7
1.2.2. Rutherford's model of the atom	9
1.2.3. Bohr's model of the hydrogen atom	10
1.2.4. Multielectron atoms	12
1.2.5. Nuclear structure	14
1.2.6. Nuclear reactions	15
1.2.7. Radioactivity	16
1.2.8. Activation of nuclides	19
1.2.9. Modes of radioactive decay	20
1.3. ELECTRON INTERACTIONS	22
1.3.1. Electron-orbital electron interactions	23
1.3.2. Electron-nucleus interactions	23
1.3.3. Stopping power	24
1.3.4. Mass scattering power	25
1.4. PHOTON INTERACTIONS	26
1.4.1. Types of indirectly ionizing photon radiation	26
1.4.2. Photon beam attenuation	26
1.4.3. Types of photon interaction	28
1.4.4. Photoelectric effect	28
1.4.5. Coherent (Rayleigh) scattering	29

1.4.6.	Compton effect (incoherent scattering)	30	
1.4.7.	Pair production	32	
1.4.8.	Photonuclear reactions	34	
1.4.9.	Contributions to attenuation coefficients	34	
1.4.10.	Relative predominance of individual effects	36	
1.4.11.	Effects following photon interactions	37	
1.4.12.	Summary of photon interactions	38	
1.4.13.	Example of photon attenuation	40	
1.4.14.	Production of vacancies in atomic shells	41	
BIBLIOGRAPHY.....		43	
CHAPTER 2. DOSIMETRIC PRINCIPLES, QUANTITIES AND UNITS			45
2.1.	INTRODUCTION	45	
2.2.	PHOTON FLUENCE AND ENERGY FLUENCE	45	
2.3.	KERMA	48	
2.4.	CEMA	48	
2.5.	ABSORBED DOSE	49	
2.6.	STOPPING POWER	49	
2.7.	RELATIONSHIPS BETWEEN VARIOUS DOSIMETRIC QUANTITIES	54	
2.7.1.	Energy fluence and kerma (photons)	54	
2.7.2.	Fluence and dose (electrons)	56	
2.7.3.	Kerma and dose (charged particle equilibrium)	57	
2.7.4.	Collision kerma and exposure	60	
2.8.	CAVITY THEORY	61	
2.8.1.	Bragg-Gray cavity theory	61	
2.8.2.	Spencer-Attix cavity theory	62	
2.8.3.	Considerations in the application of cavity theory to ionization chamber calibration and dosimetry protocols .	64	
2.8.4.	Large cavities in photon beams	66	
2.8.5.	Burlin cavity theory for photon beams	66	
2.8.6.	Stopping power ratios	68	
BIBLIOGRAPHY.....		70	

CHAPTER 3. RADIATION

3.1.	INTRODUCTION ...	
3.2.	PROPERTIES OF DOSE	
3.2.1.	Accuracy and precision	
3.2.1.1.	Type I	
3.2.1.2.	Type II	
3.2.1.3.	Combined	
3.2.2.	Linearity	
3.2.3.	Dose rate dependence	
3.2.4.	Energy dependence	
3.2.5.	Directional dependence	
3.2.6.	Spatial resolution	
3.2.7.	Readout conversion	
3.2.8.	Convenience of use	
3.3.	IONIZATION CHAMBERS	
3.3.1.	Chambers and detectors	
3.3.2.	Cylindrical (thimble)	
3.3.3.	Parallel-plate (Gerson-Rogers)	
3.3.4.	Brachytherapy	
3.3.5.	Extrapolation chamber	
3.4.	FILM DOSIMETRY	
3.4.1.	Radiographic film	
3.4.2.	Radiochromic	
3.5.	LUMINESCENCE DOSIMETRY	
3.5.1.	Thermoluminescence	
3.5.2.	Thermoluminescence	
3.5.3.	Optically stimulated luminescence	
3.6.	SEMICONDUCTOR DOSIMETRY	
3.6.1.	Silicon diode detectors	
3.6.2.	MOSFET dosimeters	
3.7.	OTHER DOSIMETRY	
3.7.1.	Alanine/electron spin resonance	
	system	
3.7.2.	Plastic scintillation	
3.7.3.	Diamond dosimetry	

CHAPTER 3. RADIATION DOSIMETERS	71
3.1. INTRODUCTION	71
3.2. PROPERTIES OF DOSIMETERS	72
3.2.1. Accuracy and precision	72
3.2.1.1. Type A standard uncertainties	72
3.2.1.2. Type B standard uncertainties	73
3.2.1.3. Combined and expanded uncertainties	73
3.2.2. Linearity	74
3.2.3. Dose rate dependence	74
3.2.4. Energy dependence	75
3.2.5. Directional dependence	76
3.2.6. Spatial resolution and physical size	76
3.2.7. Readout convenience	76
3.2.8. Convenience of use	76
3.3. IONIZATION CHAMBER DOSIMETRY SYSTEMS	77
3.3.1. Chambers and electrometers	77
3.3.2. Cylindrical (thimble type) ionization chambers	78
3.3.3. Parallel-plate (plane-parallel) ionization chambers	79
3.3.4. Brachytherapy chambers	79
3.3.5. Extrapolation chambers	79
3.4. FILM DOSIMETRY	81
3.4.1. Radiographic film	81
3.4.2. Radiochromic film	84
3.5. LUMINESCENCE DOSIMETRY	84
3.5.1. Thermoluminescence	85
3.5.2. Thermoluminescent dosimeter systems	86
3.5.3. Optically stimulated luminescence systems	88
3.6. SEMICONDUCTOR DOSIMETRY	89
3.6.1. Silicon diode dosimetry systems	89
3.6.2. MOSFET dosimetry systems	90
3.7. OTHER DOSIMETRY SYSTEMS	91
3.7.1. Alanine/electron paramagnetic resonance dosimetry system	91
3.7.2. Plastic scintillator dosimetry system	92
3.7.3. Diamond dosimeters	92

3.7.4.	Gel dosimetry systems	93
3.8.	PRIMARY STANDARDS	94
3.8.1.	Primary standard for air kerma in air	95
3.8.2.	Primary standards for absorbed dose to water	95
3.8.3.	Ionometric standard for absorbed dose to water	96
3.8.4.	Chemical dosimetry standard for absorbed dose to water	96
3.8.5.	Calorimetric standard for absorbed dose to water	97
3.9.	SUMMARY OF SOME COMMONLY USED DOSIMETRIC SYSTEMS	97
	BIBLIOGRAPHY.....	99
CHAPTER 4. RADIATION MONITORING INSTRUMENTS		101
4.1.	INTRODUCTION	101
4.2.	OPERATIONAL QUANTITIES FOR RADIATION MONITORING	102
4.3.	AREA SURVEY METERS	103
4.3.1.	Ionization chambers	105
4.3.2.	Proportional counters	105
4.3.3.	Neutron area survey meters	105
4.3.4.	Geiger-Müller counters	106
4.3.5.	Scintillator detectors	107
4.3.6.	Semiconductor detectors	107
4.3.7.	Commonly available features of area survey meters	108
4.3.8.	Calibration of survey meters	108
4.3.9.	Properties of survey meters	110
4.3.9.1.	Sensitivity	110
4.3.9.2.	Energy dependence	110
4.3.9.3.	Directional dependence	111
4.3.9.4.	Dose equivalent range	111
4.3.9.5.	Response time	111
4.3.9.6.	Overload characteristics	111
4.3.9.7.	Long term stability	112
4.3.9.8.	Discrimination between different types of radiation	112
4.3.9.9.	Uncertainties in area survey measurements ...	112
4.4.	INDIVIDUAL MONITORING	113
4.4.1.	Film badge	113

4.4.2.	Thermoluminescence
4.4.3.	Radiophotography
4.4.4.	Optically stimulated luminescence
4.4.5.	Direct reading dosimeters
4.4.6.	Calibration
4.4.7.	Properties of dosimeters
4.4.7.1.	Sensitivity
4.4.7.2.	Energy dependence
4.4.7.3.	Directional dependence
4.4.7.4.	Dose equivalent range
4.4.7.5.	Response time
4.4.7.6.	Overload characteristics

BIBLIOGRAPHY

CHAPTER 5. TREATMENT OF BEAM RADIATION

5.1.	INTRODUCTION
5.2.	X RAY BEAMS AND THERAPY
5.2.1.	Characteristics of X ray beams
5.2.2.	Bremsstrahlung
5.2.3.	X ray target
5.2.4.	Clinical X ray beams
5.2.5.	X ray beam modifiers
5.2.6.	X ray machine
5.3.	GAMMA RAY BEAMS
5.3.1.	Basic properties
5.3.2.	Teletherapy
5.3.3.	Teletherapy
5.3.4.	Teletherapy
5.3.5.	Dose delivery
5.3.6.	Collimator
5.4.	PARTICLE ACCELERATORS
5.4.1.	Betatron
5.4.2.	Cyclotron
5.4.3.	Microtron

4.4.2.	Thermoluminescence dosimetry badge	115
4.4.3.	Radiophotoluminescent glass dosimetry systems	116
4.4.4.	Optically stimulated luminescence systems	116
4.4.5.	Direct reading personal monitors	117
4.4.6.	Calibration of personal dosimeters	118
4.4.7.	Properties of personal monitors	118
4.4.7.1.	Sensitivity	118
4.4.7.2.	Energy dependence	119
4.4.7.3.	Uncertainties in personal monitoring measurements	119
4.4.7.4.	Equivalent dose range	119
4.4.7.5.	Directional dependence	120
4.4.7.6.	Discrimination between different types of radiation	120

BIBLIOGRAPHY.....	120
-------------------	-----

CHAPTER 5. TREATMENT MACHINES FOR EXTERNAL BEAM RADIOTHERAPY 123

5.1.	INTRODUCTION	123
5.2.	X RAY BEAMS AND X RAY UNITS	124
5.2.1.	Characteristic X rays	124
5.2.2.	Bremsstrahlung (continuous) X rays	124
5.2.3.	X ray targets	125
5.2.4.	Clinical X ray beams	126
5.2.5.	X ray beam quality specifiers	127
5.2.6.	X ray machines for radiotherapy	127
5.3.	GAMMA RAY BEAMS AND GAMMA RAY UNITS	129
5.3.1.	Basic properties of gamma rays	129
5.3.2.	Teletherapy machines	130
5.3.3.	Teletherapy sources	130
5.3.4.	Teletherapy source housing	131
5.3.5.	Dose delivery with teletherapy machines	132
5.3.6.	Collimator and penumbra	132
5.4.	PARTICLE ACCELERATORS	132
5.4.1.	Betatron	134
5.4.2.	Cyclotron	134
5.4.3.	Microtron	135

5.5.	LINACS	136
5.5.1.	Linac generations	137
5.5.2.	Safety of linac installations	137
5.5.3.	Components of modern linacs	138
5.5.4.	Configuration of modern linacs	138
5.5.5.	Injection system	140
5.5.6.	Radiofrequency power generation system	143
5.5.7.	Accelerating waveguide	143
5.5.8.	Microwave power transmission	144
5.5.9.	Auxiliary system	145
5.5.10.	Electron beam transport	146
5.5.11.	Linac treatment head	146
5.5.12.	Production of clinical photon beams in a linac	147
5.5.13.	Beam collimation	148
5.5.14.	Production of clinical electron beams in a linac	149
5.5.15.	Dose monitoring system	149
5.6.	RADIOTHERAPY WITH PROTONS, NEUTRONS AND HEAVY IONS	151
5.7.	SHIELDING CONSIDERATIONS	152
5.8.	COBALT-60 TELETHERAPY UNITS VERSUS LINACS	153
5.9.	SIMULATORS AND COMPUTED TOMOGRAPHY SIMULATORS	156
5.9.1.	Radiotherapy simulator	157
5.9.2.	Computed tomography simulator	158
5.10.	TRAINING REQUIREMENTS	159
	BIBLIOGRAPHY	160
CHAPTER 6. EXTERNAL PHOTON BEAMS: PHYSICAL ASPECTS		161
6.1.	INTRODUCTION	161
6.2.	QUANTITIES USED IN DESCRIBING A PHOTON BEAM ..	161
6.2.1.	Photon fluence and photon fluence rate	162
6.2.2.	Energy fluence and energy fluence rate	162
6.2.3.	Air kerma in air	163
6.2.4.	Exposure in air	164
6.2.5.	Dose to small mass of medium in air	164
6.3.	PHOTON BEAM SOURCES	166

6.4.	INVERSE SQUARE
6.5.	PENETRATION OF PHANTOM OR PA
6.5.1.	Surface dose
6.5.2.	Buildup region
6.5.3.	Depth of dose
6.5.4.	Exit dose
6.6.	RADIATION TREATMENT
6.6.1.	Radiation beam
6.6.2.	Collimator
6.6.3.	Peak scatter
6.6.4.	Relative dose
6.7.	CENTRAL AXIS D SOURCE TO SURF
6.7.1.	Percentage
6.7.2.	Scatter function
6.8.	CENTRAL AXIS D DISTANCE SET-U
6.8.1.	Tissue-air
6.8.2.	Relationship PDD(d, A)
6.8.3.	Scatter-air
6.8.4.	Relationship
6.8.5.	Tissue-ph
6.8.6.	Relationship PDD(z, A)
6.8.7.	Scatter-ma
6.9.	OFF-AXIS RATIO
6.9.1.	Beam flatness
6.9.2.	Beam symmetry
6.10.	ISODOSE DISTRIBUTION
6.11.	SINGLE FIELD IS
6.11.1.	Correction for beam incidence
6.11.1.1.	Beam incidence
6.11.1.2.	Beam incidence

6.4.	INVERSE SQUARE LAW	167
6.5.	PENETRATION OF PHOTON BEAMS INTO A PHANTOM OR PATIENT	169
6.5.1.	Surface dose	171
6.5.2.	Buildup region	171
6.5.3.	Depth of dose maximum z_{\max}	172
6.5.4.	Exit dose	172
6.6.	RADIATION TREATMENT PARAMETERS	172
6.6.1.	Radiation beam field size	173
6.6.2.	Collimator factor	174
6.6.3.	Peak scatter factor	175
6.6.4.	Relative dose factor	177
6.7.	CENTRAL AXIS DEPTH DOSES IN WATER: SOURCE TO SURFACE DISTANCE SET-UP	179
6.7.1.	Percentage depth dose	179
6.7.2.	Scatter function	181
6.8.	CENTRAL AXIS DEPTH DOSES IN WATER: SOURCE TO AXIS DISTANCE SET-UP	183
6.8.1.	Tissue-air ratio	184
6.8.2.	Relationship between $TAR(d, A_Q, hv)$ and $PDD(d, A, f, hv)$	185
6.8.3.	Scatter-air ratio	189
6.8.4.	Relationship between $SAR(d, A_Q, hv)$ and $S(z, A, f, hv)$..	190
6.8.5.	Tissue-phantom ratio and tissue-maximum ratio	190
6.8.6.	Relationship between $TMR(z, A_Q, hv)$ and $PDD(z, A, f, hv)$	192
6.8.7.	Scatter-maximum ratio	193
6.9.	OFF-AXIS RATIOS AND BEAM PROFILES	194
6.9.1.	Beam flatness	196
6.9.2.	Beam symmetry	197
6.10.	ISODOSE DISTRIBUTIONS IN WATER PHANTOMS	197
6.11.	SINGLE FIELD ISODOSE DISTRIBUTIONS IN PATIENTS ..	199
6.11.1.	Corrections for irregular contours and oblique beam incidence	200
6.11.1.1.	Effective source to surface distance method ...	201
6.11.1.2.	Tissue-air ratio or tissue-maximum ratio method	202

6.11.1.3.	Isodose shift method	202
6.11.2.	Missing tissue compensation	202
6.11.2.1.	Wedge filters	203
6.11.2.2.	Bolus	203
6.11.2.3.	Compensators	203
6.11.3.	Corrections for tissue inhomogeneities	204
6.11.4.	Model based algorithms	205
6.12.	CLARKSON SEGMENTAL INTEGRATION	206
6.13.	RELATIVE DOSE MEASUREMENTS WITH IONIZATION CHAMBERS	209
6.14.	DELIVERY OF DOSE WITH A SINGLE EXTERNAL BEAM	212
6.15.	EXAMPLE OF DOSE CALCULATION	213
6.16.	SHUTTER CORRECTION TIME	215
	BIBLIOGRAPHY.....	216
 CHAPTER 7. CLINICAL TREATMENT PLANNING IN EXTERNAL PHOTON BEAM RADIOTHERAPY		
		219
7.1.	INTRODUCTION	219
7.2.	VOLUME DEFINITION	219
7.2.1.	Gross tumour volume	220
7.2.2.	Clinical target volume	220
7.2.3.	Internal target volume	221
7.2.4.	Planning target volume	221
7.2.5.	Organ at risk	222
7.3.	DOSE SPECIFICATION	222
7.4.	PATIENT DATA ACQUISITION AND SIMULATION	223
7.4.1.	Need for patient data	223
7.4.2.	Nature of patient data	223
7.4.2.1.	Two dimensional treatment planning	223
7.4.2.2.	Three dimensional treatment planning	224
7.4.3.	Treatment simulation	225
7.4.4.	Patient treatment position and immobilization devices ..	226
7.4.5.	Patient data requirements	228
7.4.6.	Conventional treatment simulation	229
7.4.6.1.	Simulators	229

7.4.6.2.	Lo
	org
7.4.6.3.	De
7.4.6.4.	Ac
7.4.7.	Computed to
	treatment sin
7.4.7.1.	Co
	acc
7.4.7.2.	De
	ge
7.4.8.	Computed to
7.4.8.1.	Co
7.4.8.2.	Vir
7.4.8.3.	Di
7.4.8.4.	Be
7.4.8.5.	Vir
7.4.9.	Conventio
	simulator ..
7.4.10.	Magnetic res
7.4.11.	Summary of
 7.5. CLINICAL CONSID	
7.5.1.	Isodose curv
7.5.2.	Wedge filters
7.5.3.	Bolus
7.5.4.	Compensatin
7.5.5.	Corrections f
7.5.5.1.	Iso
7.5.5.2.	Eff
7.5.5.3.	Tis
7.5.6.	Corrections f
7.5.6.1.	Tis
7.5.6.2.	Bat
7.5.6.3.	Eq
7.5.6.4.	Iso
7.5.7.	Beam combin
7.5.7.1.	We
7.5.7.2.	Fix
	tech
7.5.7.3.	Par
7.5.7.4.	Mu

7.4.6.2.	Localization of the target volume and organs at risk	230
7.4.6.3.	Determination of the treatment beam geometry	230
7.4.6.4.	Acquisition of patient data	230
7.4.7.	Computed tomography based conventional treatment simulation	230
7.4.7.1.	Computed tomography based patient data acquisition	230
7.4.7.2.	Determination of the treatment beam geometry	232
7.4.8.	Computed tomography based virtual simulation	233
7.4.8.1.	Computed tomography simulator	233
7.4.8.2.	Virtual simulation	233
7.4.8.3.	Digitally reconstructed radiographs	234
7.4.8.4.	Beam's eye view	234
7.4.8.5.	Virtual simulation procedure	235
7.4.9.	Conventional simulator versus computed tomography simulator	237
7.4.10.	Magnetic resonance imaging for treatment planning	238
7.4.11.	Summary of simulation procedures	240
7.5.	CLINICAL CONSIDERATIONS FOR PHOTON BEAMS	241
7.5.1.	Isodose curves	241
7.5.2.	Wedge filters	241
7.5.3.	Bolus	244
7.5.4.	Compensating filters	245
7.5.5.	Corrections for contour irregularities	246
7.5.5.1.	Isodose shift method	246
7.5.5.2.	Effective attenuation coefficient method	248
7.5.5.3.	Tissue-air ratio method	248
7.5.6.	Corrections for tissue inhomogeneities	248
7.5.6.1.	Tissue-air ratio method	249
7.5.6.2.	Batho power law method	250
7.5.6.3.	Equivalent tissue-air ratio method	250
7.5.6.4.	Isodose shift method	250
7.5.7.	Beam combinations and clinical application	251
7.5.7.1.	Weighting and normalization	251
7.5.7.2.	Fixed source to surface distance versus isocentric techniques	251
7.5.7.3.	Parallel opposed beams	252
7.5.7.4.	Multiple coplanar beams	253

7.5.7.5.	Rotational techniques	254
7.5.7.6.	Multiple non-coplanar beams	255
7.5.7.7.	Field matching	255
7.6.	TREATMENT PLAN EVALUATION	256
7.6.1.	Isodose curves	257
7.6.2.	Orthogonal planes and isodose surfaces	257
7.6.3.	Dose statistics	257
7.6.4.	Dose-volume histograms	258
7.6.4.1.	Direct dose-volume histogram	259
7.6.4.2.	Cumulative dose-volume histogram	259
7.6.5.	Treatment evaluation	260
7.6.5.1.	Port films	261
7.6.5.2.	On-line portal imaging	262
7.7.	TREATMENT TIME AND MONITOR UNIT CALCULATIONS	264
7.7.1.	Treatment time and monitor unit calculations for a fixed source to surface distance set-up	265
7.7.2.	Monitor unit and treatment time calculations for isocentric set-ups	267
7.7.3.	Normalization of dose distributions	270
7.7.4.	Inclusion of output parameters in the dose distribution	270
7.7.5.	Treatment time calculation for orthovoltage and cobalt-60 units	271
	BIBLIOGRAPHY.....	271
CHAPTER 8. ELECTRON BEAMS: PHYSICAL AND CLINICAL ASPECTS		
8.1.	CENTRAL AXIS DEPTH DOSE DISTRIBUTIONS IN WATER	273
8.1.1.	General shape of the depth dose curve	273
8.1.2.	Electron interactions with an absorbing medium	274
8.1.3.	Inverse square law (virtual source position)	276
8.1.4.	Range concept	277
8.1.5.	Buildup region (depths between the surface and z_{\max} (i.e. $0 \leq z \leq z_{\max}$))	279
8.1.6.	Dose distribution beyond z_{\max} ($z > z_{\max}$)	279

8.2.	DOSIMETRIC PARAMETERS	
8.2.1.	Electron beam	
8.2.2.	Typical depth dose	
8.2.3.	Percentage depth dose	
8.2.3.1.	Percentage depth dose at depth d in field	
8.2.3.2.	Percentage depth dose at depth d in incidence	
8.2.4.	Output factor	
8.2.5.	Therapeutic ratio	
8.2.6.	Profiles and output	
8.2.7.	Flatness and symmetry	
8.3.	CLINICAL CONSIDERATIONS IN BEAM THERAPY	
8.3.1.	Dose specification	
8.3.2.	Small field size	
8.3.3.	Isodose curves	
8.3.4.	Field shaping	
8.3.4.1.	Electron beam	
8.3.4.2.	Shielding	
8.3.4.3.	Interlocks	
8.3.4.4.	External beam	
8.3.5.	Irregular surface	
8.3.6.	Bolus	
8.3.7.	Inhomogeneity	
8.3.7.1.	Coefficient of variation	
8.3.7.2.	Scatter	
8.3.8.	Electron beam	
8.3.8.1.	Material	
8.3.8.2.	Material	
8.3.9.	Electron arc therapy	
8.3.10.	Electron therapy	

BIBLIOGRAPHY.....

CHAPTER 9. CALIBRATION OF BEAMS

9.1. INTRODUCTION

8.2.	DOSIMETRIC PARAMETERS OF ELECTRON BEAMS	281
8.2.1.	Electron beam energy specification	281
8.2.2.	Typical depth dose parameters as a function of energy . .	281
8.2.3.	Percentage depth dose	282
8.2.3.1.	Percentage depth doses for small electron field sizes	282
8.2.3.2.	Percentage depth doses for oblique beam incidence	283
8.2.4.	Output factors	284
8.2.5.	Therapeutic range R_{90}	285
8.2.6.	Profiles and off-axis ratios	285
8.2.7.	Flatness and symmetry	285

8.3.	CLINICAL CONSIDERATIONS IN ELECTRON BEAM THERAPY	286
8.3.1.	Dose specification and reporting	286
8.3.2.	Small field sizes	287
8.3.3.	Isodose curves	287
8.3.4.	Field shaping	289
8.3.4.1.	Electron applicators	289
8.3.4.2.	Shielding and cut-outs	289
8.3.4.3.	Internal shielding	290
8.3.4.4.	Extended source to surface distance treatments	290
8.3.5.	Irregular surface correction	291
8.3.6.	Bolus	291
8.3.7.	Inhomogeneity corrections	292
8.3.7.1.	Coefficient of equivalent thickness	292
8.3.7.2.	Scatter perturbation (edge) effects	293
8.3.8.	Electron beam combinations	295
8.3.8.1.	Matched (abutted) electron fields	295
8.3.8.2.	Matched photon and electron fields	295
8.3.9.	Electron arc therapy	295
8.3.10.	Electron therapy treatment planning	298
	BIBLIOGRAPHY	299

CHAPTER 9.	CALIBRATION OF PHOTON AND ELECTRON BEAMS	301
------------	---	-----

9.1.	INTRODUCTION	301
------	------------------------	-----

9.1.1.	Calorimetry	302
9.1.2.	Fricke dosimetry	303
9.1.3.	Ionization chamber dosimetry	304
9.1.4.	Mean energy expended in air per ion pair formed	304
9.1.5.	Reference dosimetry with ionization chambers	305
9.1.5.1.	Standard free air ionization chambers	305
9.1.5.2.	Cavity ionization chambers	306
9.1.5.3.	Phantom embedded extrapolation chambers ..	306
9.1.6.	Clinical beam calibration and measurement chain	307
9.1.7.	Dosimetry protocols	307
9.2.	IONIZATION CHAMBER BASED DOSIMETRY SYSTEMS ..	308
9.2.1.	Ionization chambers	308
9.2.2.	Electrometer and power supply	309
9.2.3.	Phantoms	310
9.3.	CHAMBER SIGNAL CORRECTION FOR INFLUENCE QUANTITIES	312
9.3.1.	Air temperature, pressure and humidity effects: $k_{T,P}$	312
9.3.2.	Chamber polarity effects: polarity correction factor k_{pol}	313
9.3.3.	Chamber voltage effects: recombination correction factor k_{sat}	314
9.3.4.	Chamber leakage currents	318
9.3.5.	Chamber stem effects	319
9.4.	DETERMINATION OF ABSORBED DOSE USING CALIBRATED IONIZATION CHAMBERS	319
9.4.1.	Air kerma based protocols	320
9.4.2.	Absorbed dose to water based protocols	323
9.5.	STOPPING POWER RATIOS	326
9.5.1.	Stopping power ratios for electron beams	326
9.5.2.	Stopping power ratios for photon beams	327
9.6.	MASS-ENERGY ABSORPTION COEFFICIENT RATIOS ...	328
9.7.	PERTURBATION CORRECTION FACTORS	329
9.7.1.	Displacement perturbation factor p_{dis} and effective point of measurement	330
9.7.2.	Chamber wall perturbation factor p_{wall}	331

9.7.3.	Central elect
9.7.4.	Cavity or flu
9.8.	BEAM QUALITY SI
9.8.1.	Beam quality photon beam
9.8.2.	Beam quality photon beam
9.8.3.	Beam quality electron beam
9.9.	CALIBRATION OF AND ELECTRON B
9.9.1.	Calibration o kerma in air
9.9.2.	Calibration o the dose to w
9.9.3.	Calibration o air kerma in a
9.9.4.	Calibration o dose to water
9.10.	KILOVOLTAGE DO
9.10.1.	Specific featu
9.10.2.	Air kerma ba (medium ene
9.10.3.	Air kerma ba photon energ
9.10.4.	Air kerma in low energies
9.10.5.	Absorbed dos
9.11.	ERROR AND UNCE CHAMBER MEASU
9.11.1.	Errors and un
9.11.2.	Classification
9.11.3.	Uncertainties

BIBLIOGRAPHY...

9.7.3.	Central electrode perturbation p_{cel}	333
9.7.4.	Cavity or fluence perturbation correction p_{cav}	334
9.8.	BEAM QUALITY SPECIFICATION	335
9.8.1.	Beam quality specification for kilovoltage photon beams	336
9.8.2.	Beam quality specification for megavoltage photon beams	337
9.8.3.	Beam quality specification for megavoltage electron beams	339
9.9.	CALIBRATION OF MEGAVOLTAGE PHOTON AND ELECTRON BEAMS: PRACTICAL ASPECTS	342
9.9.1.	Calibration of megavoltage photon beams based on the air kerma in air calibration coefficient $N_{\text{K,Co}}$	342
9.9.2.	Calibration of megavoltage photon beams based on the dose to water calibration coefficient $N_{\text{D,w,Co}}$	343
9.9.3.	Calibration of megavoltage electron beams based on the air kerma in air calibration coefficient $N_{\text{K,Co}}$	345
9.9.4.	Calibration of high energy electron beams based on the dose to water calibration coefficient $N_{\text{D,w,Co}}$	346
9.10.	KILOVOLTAGE DOSIMETRY	347
9.10.1.	Specific features of kilovoltage beams	347
9.10.2.	Air kerma based in-phantom calibration method (medium energies)	348
9.10.3.	Air kerma based backscatter method (low and medium photon energies)	349
9.10.4.	Air kerma in air based calibration method for very low energies	351
9.10.5.	Absorbed dose to water based calibration method	351
9.11.	ERROR AND UNCERTAINTY ANALYSIS FOR IONIZATION CHAMBER MEASUREMENTS	352
9.11.1.	Errors and uncertainties	352
9.11.2.	Classification of uncertainties	352
9.11.3.	Uncertainties in the calibration chain	352
	BIBLIOGRAPHY	353

CHAPTER 10. ACCEPTANCE TESTS AND COMMISSIONING MEASUREMENTS	355
10.1. INTRODUCTION	355
10.2. MEASUREMENT EQUIPMENT	355
10.2.1. Radiation survey equipment	355
10.2.2. Ionometric dosimetry equipment	356
10.2.3. Film	356
10.2.4. Diodes	356
10.2.5. Phantoms	357
10.2.5.1. Radiation field analyser and water phantom ..	357
10.2.5.2. Plastic phantoms	357
10.3. ACCEPTANCE TESTS	358
10.3.1. Safety checks	359
10.3.1.1. Interlocks, warning lights and patient monitoring equipment	359
10.3.1.2. Radiation survey	359
10.3.1.3. Collimator and head leakage	360
10.3.2. Mechanical checks	361
10.3.2.1. Collimator axis of rotation	361
10.3.2.2. Photon collimator jaw motion	361
10.3.2.3. Congruence of light and radiation field	362
10.3.2.4. Gantry axis of rotation	363
10.3.2.5. Patient treatment table axis of rotation	363
10.3.2.6. Radiation isocentre	364
10.3.2.7. Optical distance indicator	364
10.3.2.8. Gantry angle indicators	365
10.3.2.9. Collimator field size indicators	365
10.3.2.10. Patient treatment table motions	365
10.3.3. Dosimetry measurements	365
10.3.3.1. Photon energy	366
10.3.3.2. Photon beam uniformity	366
10.3.3.3. Photon penumbra	366
10.3.3.4. Electron energy	367
10.3.3.5. Electron beam bremsstrahlung contamination ..	367
10.3.3.6. Electron beam uniformity	368
10.3.3.7. Electron penumbra	368
10.3.3.8. Monitor characteristics	368
10.3.3.9. Arc therapy	370

10.4. COMMISSIONING	
10.4.1. Photon beam	
10.4.1.1. Cent	
10.4.1.2. Out	
10.4.1.3. Blo	
10.4.1.4. Mul	
10.4.1.5. Cen	
10.4.1.6. Dyn	
10.4.1.7. Tran	
10.4.1.8. Entr	
10.4.1.9. Virt	
10.4.2. Electron beam	
10.4.2.1. Cent	
10.4.2.2. Outp	
10.4.2.3. Tran	
10.4.2.4. Virtu	

10.5. TIME REQUIRED FOR
BIBLIOGRAPHY.....

CHAPTER 11. COMPUTER
SYSTEMS FOR
RADIOTHERAPY

11.1. INTRODUCTION	
11.2. SYSTEM HARDWARE	
11.2.1. Treatment plan	
11.2.2. Treatment plan	
11.3. SYSTEM SOFTWARE	
11.3.1. Calculation algo	
11.3.2. Beam modifiers	
11.3.2.1. Photo	
11.3.2.2. Electr	
11.3.3. Heterogeneity c	
11.3.4. Image display an	
11.3.5. Optimization an	
11.3.6. Record and veri	
11.3.7. Biological mode	

10.4. COMMISSIONING	370
10.4.1. Photon beam measurements	370
10.4.1.1. Central axis percentage depth doses	370
10.4.1.2. Output factors	371
10.4.1.3. Blocking tray factors	373
10.4.1.4. Multileaf collimators	373
10.4.1.5. Central axis wedge transmission factors	374
10.4.1.6. Dynamic wedge	375
10.4.1.7. Transverse beam profiles/off-axis energy changes	376
10.4.1.8. Entrance dose and interface dosimetry	376
10.4.1.9. Virtual source position	377
10.4.2. Electron beam measurements	378
10.4.2.1. Central axis percentage depth dose	378
10.4.2.2. Output factors	380
10.4.2.3. Transverse beam profiles	383
10.4.2.4. Virtual source position	383
10.5. TIME REQUIRED FOR COMMISSIONING	384
BIBLIOGRAPHY.....	385

CHAPTER 11. COMPUTERIZED TREATMENT PLANNING
SYSTEMS FOR EXTERNAL PHOTON BEAM
RADIOTHERAPY

11.1. INTRODUCTION	387
11.2. SYSTEM HARDWARE	388
11.2.1. Treatment planning system hardware	388
11.2.2. Treatment planning system configurations	389
11.3. SYSTEM SOFTWARE AND CALCULATION ALGORITHMS	390
11.3.1. Calculation algorithms	390
11.3.2. Beam modifiers	393
11.3.2.1. Photon beam modifiers	393
11.3.2.2. Electron beam modifiers	394
11.3.3. Heterogeneity corrections	395
11.3.4. Image display and dose-volume histograms	395
11.3.5. Optimization and monitor unit calculations	396
11.3.6. Record and verify systems	396
11.3.7. Biological modelling	397

11.4.	DATA ACQUISITION AND ENTRY	397
11.4.1.	Machine data	397
11.4.2.	Beam data acquisition and entry	398
11.4.3.	Patient data	399
11.5.	COMMISSIONING AND QUALITY ASSURANCE	400
11.5.1.	Errors	400
11.5.2.	Verification	401
11.5.3.	Spot checks	402
11.5.4.	Normalization and beam weighting	402
11.5.5.	Dose-volume histograms and optimization	403
11.5.6.	Training and documentation	403
11.5.7.	Scheduled quality assurance	403
11.6.	SPECIAL CONSIDERATIONS	404
	BIBLIOGRAPHY.....	405
CHAPTER 12. QUALITY ASSURANCE OF EXTERNAL BEAM RADIOTHERAPY		407
12.1.	INTRODUCTION	407
12.1.1.	Definitions	407
12.1.1.1.	Quality assurance	407
12.1.1.2.	Quality assurance in radiotherapy	407
12.1.1.3.	Quality control	408
12.1.1.4.	Quality standards	408
12.1.2.	Need for quality assurance in radiotherapy	408
12.1.3.	Requirements on accuracy in radiotherapy	409
12.1.4.	Accidents in radiotherapy	411
12.2.	MANAGING A QUALITY ASSURANCE PROGRAMME ...	414
12.2.1.	Multidisciplinary radiotherapy team	414
12.2.2.	Quality system/comprehensive quality assurance programme	416
12.3.	QUALITY ASSURANCE PROGRAMME FOR EQUIPMENT	418
12.3.1.	Structure of an equipment quality assurance programme	418
12.3.1.1.	Equipment specification	419
12.3.1.2.	Acceptance	419

12.3.1.3.	Com...
12.3.1.4.	Qual...
12.3.2.	Uncertainties, t...
12.3.3.	Quality assuran teletherapy ma...
12.3.4.	Quality assuran
12.3.5.	Quality assuran simulators ...
12.3.6.	Quality assuran tomography sc...
	simulation ...
12.3.7.	Quality assuran planning system
12.3.8.	Quality assuran
12.4.	TREATMENT DELIV...
12.4.1.	Patient charts
12.4.2.	Portal imaging
	12.4.2.1. Porta
	12.4.2.2. Futur
12.4.3.	In vivo dose m...
	12.4.3.1. In viv
	12.4.3.2. Use o in viv
12.4.4.	Record and ver
12.5.	QUALITY AUDIT ...
12.5.1.	Definition ...
12.5.2.	Practical qualit
	12.5.2.1. Post
	12.5.2.2. Qual
12.5.3.	What should be
BIBLIOGRAPHY....	
CHAPTER 13. BRACHYTH...	
PHYSICAL A...	
13.1.	INTRODUCTION ...
13.2.	PHOTON SOURCE CH...
13.2.1.	Practical consi

12.3.1.3.	Commissioning	420
12.3.1.4.	Quality control	420
12.3.2.	Uncertainties, tolerances and action levels	421
12.3.3.	Quality assurance programme for cobalt-60 teletherapy machines	423
12.3.4.	Quality assurance programme for linacs	425
12.3.5.	Quality assurance programme for treatment simulators	425
12.3.6.	Quality assurance programme for computed tomography scanners and computed tomography simulation	429
12.3.7.	Quality assurance programme for treatment planning systems	430
12.3.8.	Quality assurance programme for test equipment	431
12.4.	TREATMENT DELIVERY	433
12.4.1.	Patient charts	433
12.4.2.	Portal imaging	434
12.4.2.1.	Portal imaging techniques	436
12.4.2.2.	Future developments in portal imaging	439
12.4.3.	In vivo dose measurements	439
12.4.3.1.	In vivo dose measurement techniques	440
12.4.3.2.	Use of electronic portal imaging systems for in vivo dosimetry	443
12.4.4.	Record and verify systems	443
12.5.	QUALITY AUDIT	445
12.5.1.	Definition	445
12.5.2.	Practical quality audit modalities	446
12.5.2.1.	Postal audit with mailed dosimeters	446
12.5.2.2.	Quality audit visits	446
12.5.3.	What should be reviewed in a quality audit visit?	447
	BIBLIOGRAPHY	448
CHAPTER 13. BRACHYTHERAPY:		
	PHYSICAL AND CLINICAL ASPECTS	451
13.1.	INTRODUCTION	451
13.2.	PHOTON SOURCE CHARACTERISTICS	455
13.2.1.	Practical considerations	455

13.2.2.	Physical characteristics of some photon emitting brachytherapy sources	456
13.2.3.	Mechanical source characteristics	456
13.2.4.	Source specification	457
13.2.4.1.	Specification of γ ray sources	457
13.2.4.2.	Specification of β ray sources	459
13.3.	CLINICAL USE AND DOSIMETRY SYSTEMS	460
13.3.1.	Gynaecology	460
13.3.1.1.	Types of source	460
13.3.1.2.	Dose specification	460
13.3.1.3.	Source arrangement	460
13.3.1.4.	Applicators	461
13.3.1.5.	Rectal and bladder dose monitoring	461
13.3.2.	Interstitial brachytherapy	461
13.3.2.1.	Patterson-Parker system	461
13.3.2.2.	Quimby system	462
13.3.2.3.	Paris system	462
13.3.3.	Remote afterloading systems	463
13.3.4.	Permanent prostate implants	464
13.3.4.1.	Choice of radionuclide for prostate implants	465
13.3.4.2.	Planning technique: ultrasound or computed tomography	465
13.3.4.3.	Preplanning, seed placement and dose distributions	465
13.3.4.4.	Post-implant dose distributions and evaluation	465
13.3.5.	Eye plaques	466
13.3.6.	Intravascular brachytherapy	466
13.4.	DOSE SPECIFICATION AND REPORTING	467
13.4.1.	Intracavitary treatments	467
13.4.2.	Interstitial treatments	467
13.5.	DOSE DISTRIBUTIONS AROUND SOURCES	468
13.5.1.	AAPM TG 43 algorithm	468
13.5.2.	Other calculation methods for point sources	471
13.5.3.	Linear sources	473
13.5.3.1.	Unfiltered line source in air	473

13.5.3.2.	Filtered line source in air	473
13.5.3.3.	Filtered line source in air	473
13.6.	DOSE CALCULATION	
13.6.1.	Manual dose calculation	
13.6.1.1.	Manual calculation of point sources	
13.6.1.2.	Manual calculation of line sources	
13.6.2.	Computerized dose calculation	
13.6.2.1.	Source strength determination	
13.6.2.2.	Dose rate calculation	
13.6.2.3.	Dose distribution calculation	
13.6.2.4.	Optimization	
13.6.3.	Calculation of dose to target and normal tissues	
13.6.3.1.	Use of isodose lines	
13.6.3.2.	Choice of target and normal tissue points	
13.6.3.3.	Decision on the use of normal tissue points	
13.7.	COMMISSIONING OF TREATMENT PLANS	
13.7.1.	Check of the radionuclide	
13.7.2.	Check of constancy of source strength	
13.7.3.	Computer verification of a single source	
13.7.4.	Check of decay	
13.8.	SOURCE COMMISSIONING	
13.8.1.	Wipe tests	
13.8.2.	Autoradiographic	
13.8.3.	Calibration check	
13.9.	QUALITY ASSURANCE	
13.9.1.	Constancy checks	
13.9.2.	Regular checks	
13.9.2.1.	Mechanical checks	
13.9.2.2.	Source strength	
13.9.2.3.	Wipe tests	
13.9.3.	Checks of source strength	
13.9.4.	Radiation monitoring	
13.9.5.	Quality management	

13.5.3.2.	Filtered line source in air	474
13.5.3.3.	Filtered line source in water	475
13.6.	DOSE CALCULATION PROCEDURES	475
13.6.1.	Manual dose calculations	475
13.6.1.1.	Manual summation of doses	475
13.6.1.2.	Precalculated dose distributions (atlases)	475
13.6.2.	Computerized treatment planning	476
13.6.2.1.	Source localization	476
13.6.2.2.	Dose calculation	476
13.6.2.3.	Dose distribution display	476
13.6.2.4.	Optimization of dose distribution	477
13.6.3.	Calculation of treatment time	477
13.6.3.1.	Use of Patterson-Parker tables	477
13.6.3.2.	Choice of reference points	478
13.6.3.3.	Decay corrections	478
13.7.	COMMISSIONING OF BRACHYTHERAPY COMPUTER TREATMENT PLANNING SYSTEMS	479
13.7.1.	Check of the reconstruction procedure	479
13.7.2.	Check of consistency between quantities and units	479
13.7.3.	Computer versus manual dose calculation for a single source	479
13.7.4.	Check of decay corrections	479
13.8.	SOURCE COMMISSIONING	480
13.8.1.	Wipe tests	480
13.8.2.	Autoradiography and uniformity checks of activity	480
13.8.3.	Calibration chain	480
13.9.	QUALITY ASSURANCE	481
13.9.1.	Constancy check of a calibrated dosimeter	481
13.9.2.	Regular checks of sources and applicators	481
13.9.2.1.	Mechanical properties	481
13.9.2.2.	Source strength	481
13.9.2.3.	Wipe tests	482
13.9.3.	Checks of source positioning with afterloading devices ..	482
13.9.4.	Radiation monitoring around patients	482
13.9.5.	Quality management programme	482

13.10. BRACHYTHERAPY VERSUS EXTERNAL BEAM RADIOTHERAPY	483
BIBLIOGRAPHY.....	483
CHAPTER 14. BASIC RADIOBIOLOGY	485
14.1. INTRODUCTION	485
14.2. CLASSIFICATION OF RADIATIONS IN RADIOBIOLOGY ..	486
14.3. CELL CYCLE AND CELL DEATH	487
14.4. IRRADIATION OF CELLS	488
14.4.1. Direct action in cell damage by radiation	488
14.4.2. Indirect action in cell damage by radiation	488
14.4.3. Fate of irradiated cells	489
14.5. TYPE OF RADIATION DAMAGE	489
14.5.1. Timescale	489
14.5.2. Classification of radiation damage	490
14.5.3. Somatic and genetic effects	490
14.5.4. Stochastic and deterministic (non-stochastic) effects	491
14.5.5. Acute versus late tissue or organ effects	491
14.5.6. Total body radiation response	491
14.5.7. Foetal irradiation	492
14.6. CELL SURVIVAL CURVES	492
14.7. DOSE RESPONSE CURVES	494
14.8. MEASUREMENT OF RADIATION DAMAGE IN TISSUE ...	496
14.9. NORMAL AND TUMOUR CELLS: THERAPEUTIC RATIO	497
14.10. OXYGEN EFFECT	498
14.11. RELATIVE BIOLOGICAL EFFECTIVENESS	500
14.12. DOSE RATE AND FRACTIONATION	501
14.13. RADIOPROTECTORS AND RADIOSENSITIZERS	503
BIBLIOGRAPHY.....	504
CHAPTER 15. SPECIAL PROCEDURES AND TECHNIQUES IN RADIOTHERAPY	505
15.1. INTRODUCTION	505
15.2. STEREOTACTIC IRRADIATION	506
15.2.1. Physical and clinical requirements for radiosurgery	506
15.2.2. Diseases treated with stereotactic irradiation	507

15.2.3. Equipment us	
15.2.4. Historical dev	
15.2.5. Radiosurgical	
15.2.5.1. Gam	
15.2.5.2. Lina	
15.2.5.3. Mini	
15.2.6. Uncertainty in	
15.2.7. Dose prescript	
15.2.8. Commissioning	
15.2.9. Quality assura	
15.2.10. Gamma Knife	
15.2.11. Frameless ster	
15.3. TOTAL BODY IRRAD	
15.3.1. Clinical total b	
15.3.2. Diseases treat	
15.3.3. Technical aspec	
15.3.4. Total body irra	
15.3.5. Dose prescripti	
15.3.6. Commissioning	
15.3.7. Test of total bo	
15.3.8. Quality assuran	
15.4. TOTAL SKIN ELECTR	
15.4.1. Physical and cli	
irradiation ...	
15.4.2. Current total sk	
15.4.3. Selection of tot	
15.4.4. Dose calibration	
15.4.5. Skin dose rate a	
15.4.6. Commissioning	
procedure	
15.4.7. Measurement o	
dose distributio	
15.4.8. Quality assuran	
15.5. INTRAOPERATIVE R	
15.5.1. Physical and clin	
radiotherapy ..	
15.5.2. Intraoperative r	
techniques	

15.2.3.	Equipment used for stereotactic radiosurgery	507
15.2.4.	Historical development	508
15.2.5.	Radiosurgical techniques	509
15.2.5.1.	Gamma Knife	509
15.2.5.2.	Linac based radiosurgery	509
15.2.5.3.	Miniature linac on robotic arm	511
15.2.6.	Uncertainty in radiosurgical dose delivery	512
15.2.7.	Dose prescription and dose fractionation	513
15.2.8.	Commissioning of radiosurgical equipment	514
15.2.9.	Quality assurance in radiosurgery	514
15.2.10.	Gamma Knife versus linac based radiosurgery	515
15.2.11.	Frameless stereotaxy	516
15.3.	TOTAL BODY IRRADIATION	516
15.3.1.	Clinical total body irradiation categories	516
15.3.2.	Diseases treated with total body irradiation	517
15.3.3.	Technical aspects of total body irradiation	517
15.3.4.	Total body irradiation techniques	518
15.3.5.	Dose prescription point	519
15.3.6.	Commissioning of total body irradiation procedure	519
15.3.7.	Test of total body irradiation dosimetry protocol	521
15.3.8.	Quality assurance in total body irradiation	521
15.4.	TOTAL SKIN ELECTRON IRRADIATION	522
15.4.1.	Physical and clinical requirements for total skin electron irradiation	523
15.4.2.	Current total skin electron irradiation techniques	523
15.4.3.	Selection of total skin electron irradiation technique	524
15.4.4.	Dose calibration point	525
15.4.5.	Skin dose rate at the dose prescription point	525
15.4.6.	Commissioning of the total skin electron irradiation procedure	525
15.4.7.	Measurement of clinical total skin electron irradiation dose distributions	526
15.4.8.	Quality assurance in total skin electron irradiation	526
15.5.	INTRAOPERATIVE RADIOTHERAPY	527
15.5.1.	Physical and clinical requirements for intraoperative radiotherapy	527
15.5.2.	Intraoperative radiotherapy radiation modalities and techniques	527

15.5.3.	Commissioning of an intraoperative radiotherapy programme	528
15.5.4.	Quality assurance in intraoperative radiotherapy	528
15.6.	ENDOCAVITARY RECTAL IRRADIATION	529
15.6.1.	Physical and clinical requirements for endorectal irradiation	529
15.6.2.	Endorectal treatment technique	530
15.6.3.	Quality assurance in endorectal treatments	531
15.7.	CONFORMAL RADIOTHERAPY	531
15.7.1.	Basic aspects of conformal radiotherapy	531
15.7.2.	Multileaf collimators	532
15.7.3.	Acceptance testing of multileaf collimators	533
15.7.4.	Commissioning of multileaf collimators	534
15.7.5.	Quality assurance programme for multileaf collimators ..	534
15.7.6.	Intensity modulated radiotherapy	534
15.7.7.	Commissioning of intensity modulated radiotherapy systems	535
15.7.8.	Quality assurance for intensity modulated radiotherapy systems	537
15.7.9.	Dose verification for intensity modulated radiotherapy treatment plans	537
15.8.	IMAGE GUIDED RADIOTHERAPY	538
15.8.1.	Cone beam computed tomography	539
15.8.2.	Computed tomography Primatom	540
15.8.3.	Tomotherapy	541
15.8.4.	BAT system	542
15.8.5.	ExacTrac ultrasonic module	542
15.8.6.	CyberKnife	543
15.9.	ADAPTIVE RADIOTHERAPY	544
15.10.	RESPIRATORY GATED RADIOTHERAPY	544
15.11.	POSITRON EMISSION TOMOGRAPHY/COMPUTED TOMOGRAPHY SCANNERS AND POSITRON EMISSION TOMOGRAPHY/COMPUTED TOMOGRAPHY	
	IMAGE FUSION	545
	BIBLIOGRAPHY.....	548

CHAPTER 16. RADIATION PROTECTION AND RADIOISOTOPES

16.1.	INTRODUCTION
16.2.	RADIATION EFFECTS
16.2.1.	Deterministic effects
16.2.2.	Stochastic effects
16.2.3.	Effects on the environment
16.3.	INTERNATIONAL CONVENTIONS AND STANDARDS
16.4.	TYPES OF RADIATION
16.5.	QUANTITIES AND UNITS OF RADIATION PROTECTION
16.5.1.	Physical quantities
16.5.2.	Radiation protection quantities
16.5.2.1.	Organ dose
16.5.2.2.	Equivalent dose
16.5.2.3.	Effective dose
16.5.2.4.	Committed dose
16.5.2.5.	Collective dose
16.5.3.	Operational quantities
16.5.3.1.	Ambient dose equivalent
16.5.3.2.	Directional dose equivalent
16.5.3.3.	Personal dose equivalent
16.6.	BASIC FRAMEWORK
16.7.	GOVERNMENTAL REGULATIONS AND INFRASTRUCTURE
16.8.	SCOPE OF THE BASIC REGULATIONS
16.9.	RESPONSIBILITIES FOR THE IMPLEMENTATION OF BASIC SAFETY STANDARDS
16.10.	SAFETY IN THE DESIGN AND USE OF RADIATION PROTECTION EQUIPMENT
16.10.1.	Equipment
16.10.2.	Sealed sources
16.10.3.	Safety in the design and use of radiation protection equipment
16.10.3.1.	Manufacture
16.10.3.2.	Removal and disposal
	external exposure

CHAPTER 16. RADIATION PROTECTION AND SAFETY IN RADIOTHERAPY	549
16.1. INTRODUCTION	549
16.2. RADIATION EFFECTS	550
16.2.1. Deterministic effects	550
16.2.2. Stochastic effects	550
16.2.3. Effects on the embryo and foetus	551
16.3. INTERNATIONAL CONSENSUS AND RADIATION SAFETY STANDARDS	551
16.4. TYPES OF RADIATION EXPOSURE	552
16.5. QUANTITIES AND UNITS USED IN RADIATION PROTECTION	554
16.5.1. Physical quantities	554
16.5.2. Radiation protection quantities	554
16.5.2.1. Organ dose	555
16.5.2.2. Equivalent dose	555
16.5.2.3. Effective dose	556
16.5.2.4. Committed dose	557
16.5.2.5. Collective dose	558
16.5.3. Operational quantities	558
16.5.3.1. Ambient dose equivalent	558
16.5.3.2. Directional dose equivalent	558
16.5.3.3. Personal dose equivalent	559
16.6. BASIC FRAMEWORK OF RADIATION PROTECTION	559
16.7. GOVERNMENTAL REGULATION AND NATIONAL INFRASTRUCTURE	560
16.8. SCOPE OF THE BASIC SAFETY STANDARDS	561
16.9. RESPONSIBILITIES FOR IMPLEMENTATION OF BASIC SAFETY STANDARDS REQUIREMENTS	562
16.10. SAFETY IN THE DESIGN OF RADIATION SOURCES AND EQUIPMENT	562
16.10.1. Equipment	563
16.10.2. Sealed sources	565
16.10.3. Safety in the design of facilities and ancillary equipment	567
16.10.3.1. Manual brachytherapy	567
16.10.3.2. Remote control brachytherapy and external beam radiotherapy	569

16.11. SAFETY ASSOCIATED WITH ACCEPTANCE TESTS, COMMISSIONING AND OPERATION	570
16.11.1. Safe operation of external beam radiotherapy	572
16.11.2. Safe operation of brachytherapy	572
16.11.2.1. Safe operation of manual brachytherapy	574
16.11.2.2. Safe operation of remote control afterloading brachytherapy	575
16.12. SECURITY OF SOURCES	575
16.13. OCCUPATIONAL EXPOSURE	577
16.13.1. Responsibilities and conditions of service	577
16.13.2. Use of dose constraints in radiotherapy	577
16.13.3. Investigation levels for staff exposure in radiotherapy ...	578
16.13.4. Pregnant workers	578
16.13.5. Classification of areas	579
16.13.6. Local rules and supervision	579
16.13.7. Protective equipment and tools	580
16.13.8. Individual monitoring and exposure assessment	580
16.13.9. Monitoring of the workplace	581
16.13.10. Health surveillance	581
16.13.11. Records	582
16.14. MEDICAL EXPOSURE	583
16.14.1. Responsibilities for medical exposure	583
16.14.2. Justification of medical exposure	584
16.14.3. Optimization of exposure and protection	584
16.14.4. Calibration of radiotherapy sources and machines	585
16.14.5. Clinical dosimetry	587
16.14.6. Quality assurance for medical exposure	587
16.14.7. Constraints for comforters and visitors	589
16.14.8. Discharge of patients	589
16.14.9. Investigation of accidental medical exposure	590
16.15. PUBLIC EXPOSURE	591
16.15.1. Responsibilities	591
16.15.2. Access control for visitors	591
16.15.3. Radioactive waste and sources no longer in use	591
16.15.4. Monitoring of public exposure	592
16.16. POTENTIAL EXPOSURE AND EMERGENCY PLANS	592
16.16.1. Potential exposure and safety assessment	592

16.16.2. Mitigation of c	
16.16.2.1. Lost	
16.16.2.2. Stuc	
16.16.2.3. Cont	
16.16.2.4. Off-s	
16.16.2.5. Patie	
16.17. GENERAL SHIELDING	
16.17.1. Step one: Desi	
(annual dose a	
16.17.2. Step two: Calc	
(air kerma in a	
16.17.3. Step three: Att	
16.18. TYPICAL LINAC INST	
16.18.1. Workload ...	
16.18.2. Calculation of	
16.18.3. Calculation of	
16.18.4. Calculation of	
16.18.5. Determination	
16.18.6. Consideration	
energy linac ..	
16.18.7. Door of a linac	
16.18.8. Other consider	
16.19. SHIELDING DESIGN	
FACILITIES	
BIBLIOGRAPHY	
INTERNATIONAL ORGAN	
ABBREVIATIONS	
SYMBOLS	
BIBLIOGRAPHY	
INDEX	

16.16.2. Mitigation of consequences: emergency plans	593
16.16.2.1. Lost source	593
16.16.2.2. Stuck source	594
16.16.2.3. Contamination	595
16.16.2.4. Off-site accidents	595
16.16.2.5. Patient accidental exposure	595
16.17. GENERAL SHIELDING CALCULATIONS	596
16.17.1. Step one: Design dose in occupied areas (annual dose and weekly dose)	597
16.17.2. Step two: Calculation of the radiation field (air kerma in air) in the occupied area without shielding .	598
16.17.3. Step three: Attenuation by shielding barriers	599
16.18. TYPICAL LINAC INSTALLATION	600
16.18.1. Workload	600
16.18.2. Calculation of the primary barrier transmission factor ...	602
16.18.3. Calculation of the scatter barrier transmission factor ...	603
16.18.4. Calculation of the leakage barrier transmission factor ...	603
16.18.5. Determination of barrier thickness	604
16.18.6. Consideration of neutron production in a high energy linac	605
16.18.7. Door of a linac room	605
16.18.8. Other considerations	606
16.19. SHIELDING DESIGN FOR BRACHYTHERAPY	
FACILITIES	606
BIBLIOGRAPHY	607
INTERNATIONAL ORGANIZATIONS	611
ABBREVIATIONS	613
SYMBOLS	619
BIBLIOGRAPHY	627
INDEX	639