

CONTENTS

1. INTRODUCTION	1
1.1. Definitions	2
1.2. Abbreviations	2
2. PLANAR	5
2.1. Pulse height analyser	5
2.2. Uniformity	12
2.2.1. Symmetric energy window — ^{99m}Tc	12
2.2.2. Asymmetric energy windows	19
2.2.3. Different radionuclides and photon energies..	43
2.2.4. Uniformity — quantification	53
2.2.5. Uniformity — multihead systems	56
2.2.6. Corrections (linearity, energy, uniformity).....	57
2.2.7. Detector crystal	65
2.2.8. PM tube and associated electronics	82
2.2.9. Collimator	93
2.2.10. Artefacts arising from sources/phantoms	109
2.3. Spatial resolution and linearity	128
2.3.1. Intrinsic spatial resolution and linearity	128
2.3.2. System spatial resolution and linearity	144
2.4. Multiple window spatial registration	149
2.5. High count rate	156
3. SPECT	163
3.1. SPECT uniformity	163
3.2. Uniformity calibration correction	177
3.3. Centre of rotation (COR) offset (x)	182
3.4. Image alignment in γ for multihead SPECT systems	194
3.5. Detector head tilt	202
3.6. Multihead systems — other problems	207
3.7. Data acquisition and reconstruction	209
3.8. Data reconstruction — attenuation correction	214
3.9. Clinical SPECT — other examples	221
4. WHOLE BODY	235
4.1. Scan speed	235
4.2. Image size	240
4.3. Multiple scan paths	241
4.4. Detector uniformity	244
4.5. Energy window	246
4.6. Radioactive contamination	248
5. SCINTILLATION CAMERA-COMPUTER INTERFACE	251
5.1. ADC non-linearity	251
5.2. Digital image gain and offset	258
5.3. Camera to computer connection	260
5.4. ECG gate	262
5.5. Data transfer between computers	264

6.	ENVIRONMENT/RADIOACTIVITY267
6.1.	Electrical problems.....	.267
6.2.	Radioactive contamination and extraneous radiation.....	.277
7.	DISPLAY/HARD COPY283
7.1.	Film electronic formatter.....	.283
7.2.	Film processing291
7.3.	General293