

CONTENTS

E X E C U T I V E S U M M A R Y	1
Part I. The accidental overexposure	1
Part II. Medical assessment	2
Part III. Dose assessment	3
Part IV. Steps taken by the authorities in Poland following the event in Białystok	4
Part V. Conclusions, recommendations and lessons to be learned	5
Operating organization: Radiotherapy departments	5
National infrastructure for radiotherapy	5
Equipment manufacturers and suppliers	6
Medical issues	8
1. INTRODUCTION	11
1.1. Background	11
1.2. Objective	11
1.3. Scope	12
1.4. Structure..	12
2. BACKGROUND	13
2.1. The BOC	13
2.2. The clinical accelerator involved in this event	15
2.2.1. The potential for higher electron current	16
2.2.2. The dose monitoring systems and their power supply	16
2.23 International Electrotechnical Commission (IEC) requirements relevant to this event	17
2.2.4. The relevant interlock and the consequences of its failure	18
2.2.5 Control and restriction of the current in the electron gun	20
2.3. Quality assurance (QA) and the history of recent dosimetry audits at the radiotherapy department	21
2.4. Regulatory control in Poland relevant to this event	21
2.4.1. Regulatory control of medical exposures	23
2.4.2. Steps taken by the authorities in Poland following the event in Białystok	23

3.	THE ACCIDENTAL OVEREXPOSURE	24
3.1.	Description of the event	24
3.2.	The discovery of the problem	26
4.	THE RESPONSE TO THE ACCIDENTAL OVEREXPOSURE.	27
4.1.	Local actions taken on discovery of the fault condition	27
4.2.	The IAEA response and the expert mission	30
5.	DOSE ASSESSMENT	33
5.1.	Normal machine conditions	33
5.2.	Reconstruction of the fault condition	34
5.2.1.	Dose-response non-linearity of the MU counter in fault condition	35
5.2.2.	Ion collection efficiency of the ionization chambers	37
5.2.3.	Verification of the dose measured on the day of the accident	39
5.3.	Assessment of doses	41
5.3.1.	Simulation 1: Fuse break at the beginning of a treatment session	41
5.3.2.	Simulation 2: Fuse break at the beginning of a series of treatment sessions	43
5.3.3.	Simulation 3: Reverse dose reconstruction	44
5.4.	Retrospective measurements of the patient doses	45
5.4.1.	Principles of retrospective biodosimetry with FPR	45
5.4.2.	Reconstruction of the patient doses from EPR measurements	46
5.5.	Summary of findings made by the IAEA medical physics and radiation safety team	49
6.	CLINICAL COURSE OF THE OVEREXPOSURES	50
6.1.	Overview	50
6.2.	Patient 1	5 2
6.3.	Patient 2	55
6.4.	Patient 3	58
6.5.	Patient 4	6 2
6.6.	Patient 5	7 0

7.	CONCLUSIONS, RECOMMENDATIONS AND LESSONS TO BE LEARNED		76
7.1.	Operating organization: Radiotherapy departments	76	76
7.2.	National infrastructure for radiotherapy	77	77
7.3.	Equipment manufacturers and suppliers	77	77
7.4.	Medical issues	79	79
	APPENDIX I: RADIATION EFFECTS IN HUMANS		81
1.1.	Background information on radiation effects in humans	8	1
1.1.1.	Basic aspects of radiotherapy	8	2
1.1.2.	Dose-response relationships		83
1.1.3.	Fractionation of radiation exposure		84
1.2.	Radiation effects on the tissues exposed in the accident		85
1.2.1.	Effects on the skin		85
1.2.2.	Effects on breast tissue		86
1.2.3.	Effects on the lung		87
1.2.4.	Effects on the heart	8	8
1.2.5.	Effects on bone, cartilage and muscle		89
	APPENDIX II: HAEMATOLOGICAL DATA FOR THE PATIENTS		91
	REFERENCES		97
	BIBLIOGRAPHY	9	9
	EXPERTS INVOLVED IN THE ASSESSMENT OF THE ACCIDENT		101
	CONTRIBUTORS TO DRAFTING AND REVIEW		103