

CONTENTS

1.	INTRODUCTION	1
1.1.	Background	1
1.2.	Target audience	2
1.3.	Clinical use of treatment planning systems	2
1.4.	Common errors	3
1.4.1.	Education	7
1.4.2.	Verification	8
1.4.3.	Documentation	8
1.4.4.	Communication	8
1.5.	Why is quality assurance required?	8
1.5.1.	Accuracy in the radiation treatment	8
1.5.2.	Avoidance of treatment errors	9
1.6.	Brief overview of the clinical implementation of a treatment planning system	9
1.7.	Total quality management	10
1.8.	Previous reports	10
1.9.	How to use this report	11
2.	CLINICAL TREATMENT PLANNING PROCESS	12
2.1.	Radiation treatment planning process	12
2.2.	Clinical implementation of a treatment planning system	15
3.	DESCRIPTION OF RADIATION TREATMENT PLANNING SYSTEMS	16
3.1.	Hardware	16
3.2.	Software	16
3.2.1.	Three dimensional and two dimensional treatment planning systems	17
3.2.2.	Computed tomography simulation and three dimensional treatment planning systems	17
3.2.3.	Input-output	18
3.2.4.	Contouring and image display	18
3.2.5.	Beam input and calculation	19
3.2.6.	Dose display	19
3.2.7.	Plan evaluation tools	19

3.2.8. Other features	20
3.3. Single or multistation systems	20
3.4. Ancillary components	20
3.5. Third party software	21
4. ALGORITHMS USED IN RADIATION TREATMENT PLANNING	21
4.1. Introduction	21
4.2. Processing and display of anatomical data	22
4.2.1. Contours and automated segmentation	22
4.2.2. Building three dimensional objects (structures) from a series of contours	22
4.2.3. Multiplanar reconstruction and three dimensional display	25
4.2.4. Expanding three dimensional objects (structures)	25
4.2.5. Creating digital reconstructed radiographs	27
4.2.6. Registration of multiple anatomical data sets	28
4.3. Beam or source related tools	29
4.3.1. Automatic design of beam apertures	30
4.3.2. Geometrical reconstruction of sources in brachytherapy	30
4.4. Dose calculation in external beam radiotherapy	31
4.4.1. Dose calculation problem in external beam radiotherapy	31
4.4.2. Relative dose distribution	34
4.4.3. Monitor unit/time calculations and plan normalization	35
4.4.4. Other issues in dose calculation for external beam radiotherapy	36
4.5. Dose calculation in brachytherapy	40
4.5.1. Dose calculation problem in brachytherapy	40
4.5.2. Dose from point sources	43
4.5.2.1. Activity based method.....	43
4.5.2.2. TG 43 formalism.....	44
4.5.3. Dose from tubes or wires	46
4.5.4. Dose for stepping sources and optimization	46
5. QUALITY ASSESSMENT	47
5.1. Introduction	47

5.2.	Uncertainties, deviations, tolerances and errors	47
5.2.1.	Uncertainty	47
5.2.2.	Deviation	48
5.2.3.	Tolerance	49
5.2.4.	Error	49
5.3.	Quality standards, reference data, tolerances and methods of assessment for a treatment planning system	50
5.3.1.	Quality standards	50
5.3.2.	Reference data	50
5.3.3.	Tolerances for dose calculations	51
5.3.4.	Confidence limits	53
5.4.	Sources of uncertainties and limitations for a given plan	56
6.	QUALITY ASSURANCE MANAGEMENT	58
6.1.	Quality management process	58
6.2.	Treatment planning quality assurance plan	58
6.3.	Physicist responsible for the treatment planning system	58
6.4.	Personnel	61
6.5.	Communication	62
6.6.	Equipment	62
6.7.	Staff training and education	64
6.8.	Computer systems management and security	65
6.9.	Policies, procedures and documentation	66
6.10.	Common errors	67
6.10.1.	Software misinterpretation	67
6.10.2.	Normalization	68
6.10.3.	Beam parameterization	68
6.10.4.	Monitor unit/time calculations	68
6.10.5.	Inhomogeneity corrections	69
6.10.6.	Understanding software capabilities and limitations ..	69
6.10.7.	Error management	70
7.	PURCHASE PROCESS	70
7.1.	Assessment of need	71
7.2.	Request for information	73
7.3.	Vendor demonstrations, presentations and site visits	73
7.4.	Tender process: Definitions of system specifications	74
7.5.	Selection criteria	76
7.6.	Purchase	76

7.7.	Vendor and user responsibilities	77
7.7.1.	Vendor responsibilities	77
7.7.2.	User responsibilities	78
8.	ACCEPTANCE TESTING	79
8.1.	Introduction	79
8.2.	Hardware	80
8.3.	Network integration	80
8.4.	Data transfer	81
8.5.	Software	82
8.5.1.	Verifying system capabilities	82
8.5.2.	Verifying calculation capabilities	83
8.5.2.1.	Calculation on benchmark data	83
8.5.2.2.	Calculation on generic data	84
8.5.2.3.	Calculation on the institution's data	84
8.5.3.	Utility software checks	84
8.6.	Documentation	85
9.	COMMISSIONING	86
9.1.	Introduction	86
9.1.1.	Purpose	86
9.1.2.	General documentation guidelines	86
9.1.3.	General organization	87
9.1.4.	How to use Section 9	87
9.2.	System set-up and machine-source configuration	88
9.2.1.	General comments	89
9.2.2.	Computer hardware	90
9.2.3.	Computer software	91
9.2.4.	Treatment planning system configuration	91
9.2.5.	Patient database	92
9.2.6.	Treatment planning system data exchange	92
9.2.7.	Display configuration	92
9.2.8.	Planning protocols	93
9.2.9.	Computed tomography conversions	93
9.2.10.	Machine database	94
9.2.11.	Source data for brachytherapy	94
9.2.12.	Dose calculation algorithms	95
9.3.	Patient anatomical representation	96
9.3.1.	Acquisition of patient information	96

9.3.1.1.	Acquisition test 1: Manual contour acquisition	97
9.3.1.2.	Acquisition test 2: Computed tomography data acquisition.	97
9.3.2.	Entry or transfer of input anatomical data	97
9.3.2.1.	Input test 1: Digitizer calibration	97
9.3.2.2.	Input test 2: Manual contour entry	99
9.3.2.3.	Input test 3: Computed tomography data acquisition	99
9.3.2.4.	Input test 4: Computed tomography tools in the treatment planning system	100
9.3.2.5.	Input issue 1: Other imaging modalities	102
9.3.2.6.	Input issue 2: Patient database	103
9.3.3.	Anatomical model	103
9.3.3.1.	Anatomy test 1: Representation of contours without imaging	104
9.3.3.2.	Anatomy test 2: Manual contouring from computed tomography.	105
9.3.3.3.	Anatomy test 3: Automatic contouring	107
9.3.3.4.	Anatomy test 4: Editing of contours	107
9.3.3.5.	Anatomy test 5: Generating a three dimensional object description	108
9.3.3.6.	Anatomy test 6: Generating new contours (from surfaces or interpolation).	108
9.3.3.7.	Anatomy test 7: Object expansion	109
9.3.3.8.	Anatomy test 8: Creating densities for manual contours	109
9.3.3.9.	Anatomy test 9: Creating densities from computed tomography.	110
9.3.3.10.	Anatomy test 10: Creating an anatomical bolus.	110
9.3.3.11.	Anatomy test 11: Editing of CT densities	110
9.3.3.12.	Anatomy test 12: Defining points, lines and marks	110
9.3.3.13.	Anatomy test 13: Two dimensional image display	111
9.3.3.14.	Anatomy test 14: Two dimensional image display tools	111
9.3.3.15.	Anatomy test 15: Generating two dimensional reconstructed images	111

9.3.3.16.	Anatomy test 16: Three dimensional display and associated tools	112
9.3.3.17.	Anatomy test 17: Tools for the manipulation of anatomical data	112
9.3.3.18.	Anatomy test 18: Measurement tools	112
9.3.3.19.	Anatomy test 19: Basic co-ordinate system	113
9.3.3.20.	Anatomy test 20: Three dimensional co-ordinate readout	114
9.3.3.21.	Anatomy issue 1: Use of multiple image data sets and image registration	114
9.4.	External beam commissioning	115
9.4.1.	General schema for external beam commissioning	115
9.4.1.1.	Basic philosophy	115
9.4.1.2.	Methodology for dose calculation commissioning	116
9.4.1.3.	Algorithm commissioning plan	116
9.4.1.4.	Measured data	116
9.4.1.5.	Algorithm input data	118
9.4.1.6.	Calculation checks	120
9.4.1.7.	Calculation and comparison techniques	122
9.4.2.	External beam plans: Machine capabilities and beams	123
9.4.2.1.	Beam test 1: Machine description and capabilities	123
9.4.2.2.	Beam test 2: Machine readout conventions and scales	126
9.4.2.3.	Beam test 3: Machine parameter limitations	127
9.4.2.4.	Beam test 4: Collimator jaw setting	127
9.4.2.5.	Beam test 5: Asymmetric jaws	128
9.4.2.6.	Beam test 6: Block definitions and shapes	128
9.4.2.7.	Beam test 7: Multileaf collimator shaping	128
9.4.2.8.	Beam test 8: Automated field shaping	129
9.4.2.9.	Beam test 9: Beam set-up	129
9.4.2.10.	Beam test 10: Beam location	130
9.4.2.11.	Beam test 11: Gantry, collimator and table angles	130
9.4.2.12.	Beam test 12: Arcs	130
9.4.2.13.	Beam test 13: Wedges (hard, motorized and dynamic)	131
9.4.2.14.	Beam test 14: Compensators	131
9.4.2.15.	Beam test 15: Electron applicators	131

9.4.2.16.	Beam test 16: Bolus	132
9.4.2.17.	Beam test 17: Beam display on axial planes . .	132
9.4.2.18.	Beam test 18: Beam display on non-axial planes	132
9.4.2.19.	Beam test 19: Three dimensional beam displays	133
9.4.2.20.	Beam test 20: Beam's eye view display of beam and anatomy	133
9.4.2.21.	Beam test 21: Digitally reconstructed radiograph calculation and display	133
9.4.2.22.	Beam test 22: Display of portal images	134
9.4.2.23.	Beam test 23: Multiple beam isocentre functions	134
9.4.2.24.	Beam test 24: Field matching	134
9.4.2.25.	Beam test 25: Missing tissue and dose compensation	135
9.4.2.26.	Beam issue 1: Inverse planned intensity modulated radiation therapy	135
9.4.2.27.	Beam issue 2: Radiosurgery	135
9.4.2.28.	Beam issue 3: Large field techniques	136
9.4.2.29.	Beam issue 4: Complex table motions	136
9.4.3.	Photon beam commissioning	137
9.4.3.1.	Photon dose calculation issues	138
9.4.3.2.	Experimental methods for photons	138
9.4.3.3.	Tests for photons	138
9.4.3.4.	Photon test 1: Square and rectangular fields .	138
9.4.3.5.	Photon test 2: Asymmetric fields	141
9.4.3.6.	Photon test 3: Shaped fields	141
9.4.3.7.	Photon test 4: Fixed beams	142
9.4.3.8.	Photon test 5: Two dimensional arc rotations	142
9.4.3.9.	Photon test 6: Source to surface distance dependence	142
9.4.3.10.	Photon test 7: Hard wedges	143
9.4.3.11.	Photon test 8: Motorized wedge	143
9.4.3.12.	Photon test 9: Dynamic (virtual) wedge	144
9.4.3.13.	Photon test 10: Oblique incidence	145
9.4.3.14.	Photon test 11: Missing scatter	145
9.4.3.15.	Photon test 12: Buildup region behaviour . . .	146
9.4.3.16.	Photon test 13: Density correction	147
9.4.3.17.	Photon test 14: Compensation	147

9.4.3.18.	Photon test 15: Forward planned intensity modulated radiation therapy	148
9.4.3.19.	Photon issue 1: Inverse planned intensity modulated radiation therapy	148
9.4.3.20.	Photon issue 2: Radiosurgery	148
9.4.3.21.	Photon issue 3: Large field techniques	148
9.4.4.	Electron beam commissioning	148
9.4.4.1.	Issues to be studied	148
9.4.4.2.	Test procedures.	149
9.4.4.3.	Tests.	149
9.4.4.4.	Electron test 1: Square and rectangular fields.	149
9.4.4.5.	Electron test 2: Shaped fields	150
9.4.4.6.	Electron test 3: Surface collimation.	151
9.4.4.7.	Electron test 4: Source to surface distance dependence.	152
9.4.4.8.	Electron test 5: Slab bolus.	152
9.4.4.9.	Electron test 6: Shaped bolus	153
9.4.4.10.	Electron test 7: Oblique incidence.	153
9.4.4.11.	Electron test 8: Complex surface shapes.	153
9.4.4.12.	Electron test 9: Bulk density correction	154
9.4.4.13.	Electron test 10: Computed tomography based inhomogeneity corrections.	154
9.4.4.14.	Electron test 11: Arc rotations	154
9.4.5.	Operational issues	155
9.4.5.1.	Operational test 1: Algorithm choice	155
9.4.5.2.	Operational test 2: Inhomogeneity corrections.	156
9.4.5.3.	Operational test 3: Calculation validity.	157
9.4.5.4.	Operational test 4: Calculation grid and window.	157
9.4.6.	Absolute and relative dose	158
9.4.6.1.	Description of process for beam and plan normalization and monitor unit calculations	158
9.4.6.2.	Dose prescription and plan normalization issues	160
9.4.6.3.	Monitor unit and treatment time calculations	161
9.4.6.4.	MU test 1: Open fields.	161
9.4.6.5.	MU test 2: Tangential fields	163
9.4.6.6.	MU test 3: Wedged fields	163

9.4.6.7.	MU test 4: Blocked fields	163
9.4.6.8.	MU test 4a: Central axis blocked	164
9.4.6.9.	MU test 5: Multileaf collimator shaped fields	164
9.4.6.10.	MU test 5a: Central axis blocked by a multileaf collimator	165
9.4.6.11.	MU test 6: Inhomogeneity corrections	165
9.4.6.12.	MU test 7: Off-axis points	166
9.4.6.13.	MU test 8: Dose prescription	166
9.4.6.14.	MU test 9: Dose distribution units	167
9.4.6.15.	MU issue 1: Documentation for the treatment chart	167
9.4.6.16.	MU issue 2: Clinical monitor unit calculation check procedure	168
9.5.	Brachytherapy commissioning	168
9.5.1.	General schema for brachytherapy commissioning ...	168
9.5.1.1.	Basic philosophy	168
9.5.1.2.	Methodology for dose calculation commissioning	168
9.5.1.3.	Creation of the commissioning plan	169
9.5.1.4.	Reference data	170
9.5.1.5.	Dose calculation testing and comparison	170
9.5.2.	Clinical situations and tests	171
9.5.2.1.	Common techniques	171
9.5.2.2.	Algorithm related decisions	171
9.5.3.	Dose calculation tests	173
9.5.3.1.	Brachytherapy dose test 1: Source description, parameterization and reference data	173
9.5.3.2.	Brachytherapy dose test 2: Dose distribution from a single source	177
9.5.3.3.	Brachytherapy dose test 3: Dose rate for a variable length	178
9.5.3.4.	Brachytherapy dose test 4: Correction for source strength decay before application	179
9.5.3.5.	Brachytherapy dose test 5: Computation of treatment time	180
9.5.3.6.	Brachytherapy dose test 6: Correction for source strength decay during application	180
9.5.3.7.	Brachytherapy dose test 7: Dose integration for a permanent implant	181

9.5.3.8.	Brachytherapy dose test 8: Dose distribution for a stepping source	181
9.5.3.9.	Brachytherapy dose issue 1: Dose distribution around the source arrangement (and applicators)	182
9.5.3.10.	Brachytherapy dose issue 2: Changing options and editing of source characteristics	182
9.5.4.	Geometrical tests	183
9.5.4.1.	Methods for source reconstruction	183
9.5.4.2.	Tests for source reconstruction	183
9.5.4.3.	Brachytherapy geometry test 1: Quality of the geometrical reconstruction	184
9.5.4.4.	Brachytherapy geometry test 2: Source identification (manual)	186
9.5.4.5.	Brachytherapy geometry test 3: Source identification (automatic)	186
9.5.4.6.	Brachytherapy geometry test 4: Total versus active length	187
9.6.	Plan evaluation tools	187
9.6.1.	Dose display	187
9.6.1.1.	Dose display test 1: Plan normalization	188
9.6.1.2.	Dose display test 2: Isodose lines and surfaces	188
9.6.1.3.	Dose display test 3: Cold and hot spots	189
9.6.1.4.	Dose display test 4: Point dose display	189
9.6.2.	Dose-volume histograms	189
9.6.2.1.	DVH test 1: Types of dose-volume histogram	190
9.6.2.2.	DVH test 2: Plan normalization	191
9.6.2.3.	DVH test 3: Relative and absolute dose comparisons	191
9.6.2.4.	DVH test 4: Relative and absolute volume	192
9.6.2.5.	DVH test 5: Histogram dose bin size	192
9.6.2.6.	DVH test 6: Compound structures	193
9.6.2.7.	DVH test 7: Consistency with dose display	193
9.6.2.8.	DVH test 8: Calculation point sampling	194
9.6.2.9.	DVH test 9: Dose-volume histogram comparison guidelines	195
9.6.2.10.	DVH test 10: Dose and volume statistics	195
9.6.3.	Biological effects	196

9.6.3.1.	Bioeffect test 1: Normal tissue complication probabilities	196
9.6.3.2.	Bioeffect test 2: Tumour control probabilities	196
9.6.3.3.	Bioeffect test 3: Fractionation and other biological effect results	197
9.7.	Plan output and data transfer	197
9.7.1.	Plan output	197
9.7.2.	Standard plan transfer issues	199
9.7.2.1.	Transfer test 1: Treatment planning system co-ordinates and scales	200
9.7.2.2.	Transfer test 2: Co-ordinate and scale conventions for the treatment planning system and equipment	201
9.7.2.3.	Transfer test 3: Angle readings	201
9.7.2.4.	Transfer test 4: Table co-ordinates	201
9.7.2.5.	Transfer test 5: Jaws	202
9.7.2.6.	Transfer test 6: Machine definition	202
9.7.2.7.	Transfer test 7: Machine motion	202
9.7.2.8.	Transfer test 8: Wedges	203
9.7.2.9.	Transfer test 9: Blocks	203
9.7.2.10.	Transfer test 10: Multileaf collimators	204
9.7.2.11.	Transfer test 11: Electron applicators	204
9.7.2.12.	Transfer test 12: Uniqueness	204
9.7.2.13.	Transfer test 13: Miscellaneous devices	205
9.7.2.14.	Transfer test 14: Dose prescription	205
9.7.2.15.	Transfer test 15: Brachytherapy	205
9.7.2.16.	Transfer issue 1: Additional safety requirements	206
9.8.	Overall clinical tests	206
9.8.1.	Clinical test 1: Open four field box	208
9.8.2.	Clinical test 2: Same four field box, heavily corner blocked	209
9.8.3.	Clinical test 3: Wedge pair	209
9.8.4.	Clinical test 4: Computed tomography plan	209
9.8.5.	Clinical test 5: Multiple field coplanar conformal prostate plan	209
9.8.6.	Clinical test 6: Conformal non-coplanar brain plan ...	210
9.8.7.	Clinical test 7: Combined photon-electron plan	210
9.8.8.	Clinical test 8: Gynaecological application	210
9.8.9.	Clinical test 9: Two plane breast implant	210

9.8.10. Clinical test 10: Iodine-125 prostate implant	211
9.8.11. Clinical test 11: High dose rate or pulsed dose rate test case	211
10. PERIODIC QUALITY ASSURANCE	211
10.1. Introduction	211
10.2. Treatment planning system	212
10.2.1. QC test 1: Central processing unit	212
10.2.2. QC test 2: Digitizer	212
10.2.3. QC test 3: Plotter	212
10.2.4. QC test 4: Backup recovery	214
10.2.5. QC test 5: Computed tomography transfer	214
10.2.6. QC test 6: Computed tomography density and geometry	214
10.2.7. QC test 7: Patient anatomy	215
10.2.8. QC test 8: External beam revalidation	215
10.2.9. QC test 9: Monitor units/time	216
10.2.10. QC test 10: Plan details	216
10.2.11. QC test 11: Electronic plan transfer	217
10.2.12. QC test 12: Brachytherapy revalidation	217
10.2.13. QC test 13: Plan details	217
10.2.14. QC test 14: Independent dose and time check	217
10.2.15. QC test 15: Electronic plan transfer	218
10.3. Recommissioning after upgrades	218
10.3.1. Hardware	218
10.3.2. Operating system software and configuration	218
10.3.3. Treatment unit data: Recommissioning or additional data	218
10.3.4. Treatment planning system software updates and upgrades	219
11. PATIENT SPECIFIC QUALITY ASSURANCE	220
11.1. Introduction	220
11.2. Consistency during planning	220
11.3. Plan check	222
11.4. Monitor unit/time check	223
11.5. Export and handling of patient specific data: Pretreatment checks	224
11.6. Ongoing checks (weekly)	225

11.7. Other patient specific issues	225
11.8. Unusual behaviour	226
11.9. In vitro and in vivo dosimetry and imaging	226
12. SUMMARY	227
APPENDIX: TESTS TO BE PERFORMED FOR TREATMENT PLANNING SYSTEM COMMISSIONING	231
REFERENCES	249
ANNEX: PUBLISHED BENCHMARK DATA FOR ASSESSING DIFFERENT SCATTERING CONDITIONS	257
GLOSSARY	271
ABBREVIATIONS	279
CONTRIBUTORS TO DRAFTING AND REVIEW	281